



产品信息

| | |
|--------|----------------------|
| 试件名称: | 凹面 |
| 试件编号: | 1 |
| 检测人员: | 王建 |
| 日期: | 2025-12-06 10:08:51 |
| 设备序列号: | / |
| 备注: | / |
| 尺寸: | Aperture: 76.700 mm |
| 移除项: | Piston; Tilt; Power; |

Seidel Aberration

| Seidel | Value | Clock(deg) |
|-------------|------------------|------------|
| Tilt | 1.048 λ | 128.449 |
| Focus | 0.234 λ | / |
| Astigmatism | 0.040 λ | -47.784 |
| Coma | 0.024 λ | -95.094 |
| Spherical | -0.081 λ | / |

Result

| | |
|------------------|-----------------|
| PV | 0.159 λ |
| RMS | 0.018 λ |
| Power | 0.153 λ |
| Zernike_Ast | 0.020 λ |
| SAG(N) | 0.355 fr@546 |
| IRR(Δ N) | 0.368 fr@546 |