

Medical Display Glass Drawing Specification Template

Mandatory Marking Items

1. Medical-grade substrate material selection
2. Required coating stack: AR / AF / ITO composite layers
3. Overall size, thickness tolerance and safety chamfer size
4. Position of mounting holes, cable cutouts and silk screen light-shield frame
5. Mark partial masking zone if local coating restriction is required
6. Indicate daily disinfection type to match suitable medical AF coating

Special Medical Drawing Rules

All edges must adopt smooth chamfer polishing to avoid scratch risks during clinical assembly and use. We provide finished PDF engineering drawings after free DFM medical optimization to match hospital equipment frame.